


**Technical Specification clarifications in pre-bid meeting for the purchase of HPLC**

Sl. No.	Existing specifications	Clarification requested by firms	Remarks or modifications in technical specifications
1	Delivery schedule: Within 2 months from the date of issue of supply order	Delivery within 6 weeks from date of Letter of Credit or supply order is not sufficient	Letter of Credit is accepted and the minimum delivery period is 90 days
2	Sample Injector: Auto sampler should be offered. Manual sample injector should be offered for both Analytical and Semi-preparative. injection	Request to confirm only Autosampler or Manual Injector.	Autosampler is given as optional
3	Column Holder & Column: C18 Semi Preparative Column, 5um, 14 x 250 mm: 1 No	Request for Column with 10mm ID with 5micron particle or 19mm ID with 10Micron particle	15-20 mm ID with 10 Micron particle
4	PDA: Light source D2 & W Lamps	D2 Lamp or better	D2 Lamp and / or better
5	PDA: Slit width 1.2 & 8nm	Only 1.2nm or better.	Slit width 1.2 nm and/or better
6	Noise: $4.5 \times 10^{-6}$ AU	$10 \times 10^{-6}$ or better	$< 10 \times 10^{-6}$ AU
7	Drift: $4 \times 10^{-4}$ AU/hr	$10 \times 10^{-4}$ or better	$10 \times 10^{-4}$ or better
8	Flow cell pressure: 1700psi	Flow cell pressure: 1000psi or higher	Flow cell pressure:1000 to 1700psi
9	Recycle Valve:	To be removed	Recycle value is not required.
10	Easy operation with 32-bit Windows based software	Easy operation with 64-bit Windows based software	Easy operation with 64-bit Windows based software
11		Is the tender submission Online or Offline	Since it is off line tender. Hard copies of the all relevant documents to be sent as per advertisement.

  
Director (NRM)

&

Chairman of the Technical Committee

