

Corrigendum

Tender Reference No: DCPBG/DST FIST/EQUIPMENTS/PURCHASE/2021

Purchase of Dietary Fibre Analyzer

In the discussions with the representatives of the company who have attended pre bid meeting held at 5.00 PM on 28.07.2021 the following clarifications provided were approved by the technical committee

S. No.	Clarifications sought	Clarifications given
1	Scope for provision of Muffle furnace and Automated Kjeldahl system	Muffle furnace and Automated Kjeldahl system are to be provided
2	Scope for provision of Additional accessories such as Crude fibre analyser, Soxhlet apparatus (Six samples), analytical balance (0.01 mg- 50 g) and hot air oven of 150 lit capacity (Memmert type) with Stainless Steel (304 grade) up to temperature of 250°C	Additional accessories such as Crude fibre analyser, Soxhlet apparatus (Six samples), analytical balance (0.01 mg- 50 g) and hot air oven of 150 lit capacity (Memmert type) with Stainless Steel (304 grade) up to temperature of 250°C should be provided

**Director (CPBG)
Chairman
Purchase committee**