

Corrigendum

Tender Reference No: No.DCPBG/DST FIST/EQUIPMENTS/PURCHASE/2021

Purchase of Near Infra Red Analyzer

In the discussions with the representatives of the company who have attended pre bid meeting held at 5.00 PM on 28.07.2021 the following clarifications provided were approved by the technical committee

S. No.	Clarifications sought	Clarifications given
1	The machine should work in a required wavelength range of 500 -1100 nm.	The wavelength range required: 700-2000 nm
2	Measurement Mode : Transmittance	Measurement mode: Transmittance & Reflectance
3	Sample type: Grains of major crops	Sample type: Grain &Flour (Powder)

**Director (CPBG)
Chairman
Purchase committee**